

Study finds decline in performance of some COVID-19 home testing kits during omicron period

September 14 2022



Credit: Unsplash/CC0 Public Domain

The performance of some home testing kits for COVID-19 appears to have declined as the omicron variant emerged, suggests a study

published by *The BMJ* today.

The findings, based on three widely used [rapid antigen tests](#), show that performance improved when tests used both nose and throat samples compared with nasal samples only.

But only one test met the World Health Organization's standards of at least 80% [sensitivity](#) (ability to correctly identify a true positive sample) and at least 97% specificity (ability to correctly identify a true negative sample) among individuals with symptoms.

Rapid antigen tests were first introduced for use by trained professionals, but are now widely available for people to test themselves and get a result quickly, at their own convenience.

This "self-testing" approach could support early detection and self-isolation of infectious people and reduce community transmission. But there is a lack of real world evidence on the performance of nose and throat self-sampling during the omicron variant period.

To fill this knowledge gap, researchers compared the performances of three commercially available rapid antigen tests: Flowflex (Acon Laboratories), MPBio (MP Biomedicals) and Clinitest (Siemens-Healthineers), during the omicron period of the COVID-19 pandemic.

Their analysis is based on 6,497 people with COVID-19 symptoms aged 16 years or over who presented for testing at three public health service COVID-19 test sites in the Netherlands between 21 December 2021 and 10 February 2022.

All participants had a reference (PCR) test taken by a trained member of staff and were asked to complete a rapid antigen test at home as soon as possible, and within three hours of their test site visit, followed by an

online questionnaire.

Nasal self-sampling was used during the emergence of omicron, and when omicron accounted for more than 90% of infections (phase 1) and combined throat and nasal self-sampling was used when omicron accounted for more than 99% of infections (phase 2).

The researchers found that the sensitivities of the three tests performed with nasal self-sampling decreased during the emergence of omicron, from 87% to 81% for Flowflex, 83% to 76% for MPBio, and 80% to 67% for Clinitest. However, the observed decline was only statistically significant for Clinitest.

When a throat test was added to a nasal test, this improved the sensitivity of MPBio from 70% to 83% and Clinitest from 70% to 77% (but was not done for Flowflex), most notably in individuals who visited the test site for reasons other than to confirm a positive self-test result.

Only the MPBio test with combined throat and nasal self-sampling met the World Health Organization's standards for rapid antigen tests among individuals with symptoms.

These are observational findings, and the researchers point to some limitations. For example, nasal and combined throat and nasal self-sampling were conducted in different time periods, and the proportion of omicron infections may have been higher during the combined throat and nasal self-sampling period.

Also, there were slight differences in sampling methods for the reference (PCR) test, which might have influenced the results of the study.

However, the authors point out that the omicron variant was present in at

least 90% of samples in both periods and therefore do not expect that the sampling method of the reference test substantially impacted their results or their generalizability.

As such, they say, "We found that the performance of rapid antigen tests with nasal self-sampling declined during the period [omicron](#) emerged.

"We also showed that the performance of rapid antigen tests can be improved by adding oropharyngeal to nasal self-sampling. Therefore, after proper evaluation, manufacturers of rapid antigen tests should consider extending their instructions for use to include combined oropharyngeal and nasal self-sampling."

Based on these findings, they say people with COVID-19 symptoms can rely on a positive rapid antigen test result irrespective of SARS-CoV-2 variant dominance or method of self-sampling, while individuals with a negative self-test result should adhere to general preventive measures because a false negative result cannot be ruled out.

"What should we take from this study?" asks Timothy Feeney, research editor at *The BMJ*, and Charles Poole at the University of North Carolina, in an accompanying editorial.

"Firstly, that members of the general public are capable of doing their own nasal (and potentially oropharyngeal) sampling for COVID-19 testing, but the real world performance of antigen tests remains highly variable. Secondly, adding [oropharyngeal](#) testing may provide some benefit, although it is unclear how many test kits are capable of expanded use, and serial testing could be a more workable change to testing protocols."

Finally, and most importantly, are the policy implications. "In the UK and the US, policies governing use of tests to enable a return to normal

activities are confusing, poorly explained, and frequently change."

He told *The BMJ*, "Given the less than ideal performance of antigen tests, updates to guidance in the public and private sector should take this into account when suggesting action based on test results."

More information: Research: Diagnostic accuracy of covid-19 rapid antigen tests with unsupervised self-sampling in people with symptoms in the omicron period: cross sectional study Editorial: Self testing for covid-19, *The BMJ* (2022). [DOI: 10.1136/bmj-2022-071215](https://doi.org/10.1136/bmj-2022-071215)

Provided by British Medical Journal

Citation: Study finds decline in performance of some COVID-19 home testing kits during omicron period (2022, September 14) retrieved 4 July 2024 from <https://medicalxpress.com/news/2022-09-decline-covid-home-kits-omicron.html>

This document is subject to copyright. Apart from any fair dealing for the purpose of private study or research, no part may be reproduced without the written permission. The content is provided for information purposes only.